

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000112  
10:52 03/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:21  
PRESSURIZING 3:27  
STERILIZING 5:00  
136.0 C 220kPa 5:00  
136.4 C 224kPa 5:30  
136.4 C 225kPa 6:00  
136.4 C 227kPa 6:30  
136.3 C 223kPa 7:00  
136.2 C 225kPa 7:30  
136.3 C 225kPa 8:00  
136.4 C 226kPa 8:30  
136.4 C 226kPa 8:30  
VENTING 8:30  
AIR DRYING 9:32  
CYCLE COMPLETE 24:14

-----  
Digital Signature #  
D60E983C3F15A9A9  
-----