
Midmark M11 - v1.0.0
Total Cycles: 2214

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2022 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.0 F 0.0
2:00 200.7 F 4.2
4:00 200.0 F 9.2
6:00 216.2 F 7.5
8:00 228.1 F 9.7
10:00 232.6 F 8.2
12:00 241.1 F 12.4
14:00 240.1 F 10.7
16:00 253.5 F 17.3
18:00 264.8 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.5 F 29.3
2:00 271.9 F 28.7
3:00 271.7 F 28.9
4:00 272.4 F 29.3

Min 271.4 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 225.0 F 0.0
10:00 234.8 F 0.0
15:00 231.6 F 0.0
20:00 226.3 F 0.0
25:00 228.0 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:30
CYCLE COMPLETE

08/09/2022 10:58 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____