
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001650
11:58 06/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:50
STERILIZING 3:51
136.0 C 220kPa 3:51
136.1 C 224kPa 4:21
136.4 C 225kPa 4:51
136.1 C 222kPa 5:21
136.3 C 224kPa 5:51
136.4 C 225kPa 6:21
136.4 C 223kPa 6:51
136.3 C 225kPa 7:21
136.3 C 225kPa 7:21
VENTING 7:21
AIR DRYING 8:14
CYCLE COMPLETE 8:21

Digital Signature #
98345BE0FA7D529B
