
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001955
15:44 14/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:20
STERILIZING 3:26
136.5 C 226kPa 3:26
136.3 C 225kPa 3:56
136.4 C 225kPa 4:26
136.3 C 225kPa 4:56
136.4 C 226kPa 5:26
136.2 C 223kPa 5:56
136.1 C 222kPa 6:26
136.2 C 221kPa 6:56
136.2 C 221kPa 6:56
VENTING 6:56
AIR DRYING 7:48
CYCLE COMPLETE 7:57

Digital Signature #
4C3A52D0FE5DA187
