
Midmark M11 - v1.0.5
Total Cycles: 1451

Name: RITTER2
Sterilizer ID: V2313228
09/02/2022 02:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 214.3 F 0.0
2:00 206.8 F 4.3
4:00 213.9 F 10.6
6:00 228.0 F 7.6
8:00 238.0 F 11.5
10:00 243.1 F 13.0
12:00 244.2 F 12.6
14:00 257.9 F 19.9
16:00 269.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.4 F 29.5
2:00 271.5 F 28.8
3:00 272.4 F 29.7
4:00 272.0 F 29.1

Min 271.2 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 222.1 F 0.0
10:00 231.9 F 0.0
15:00 228.7 F 0.0
20:00 225.3 F 0.0
25:00 223.4 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 2:36
HEATING: 17:01
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:56:42
CYCLE COMPLETE

09/02/2022 03:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____