

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 007105  
6:24 26/03/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:43  
PRESSURIZING 3:44  
STERILIZING 4:48  
136.1 C 220kPa 4:48  
136.1 C 221kPa 5:18  
136.4 C 223kPa 5:48  
136.2 C 219kPa 6:18  
136.0 C 220kPa 6:48  
136.3 C 222kPa 7:18  
136.2 C 220kPa 7:48  
136.1 C 220kPa 8:18  
136.1 C 220kPa 8:18  
VENTING 8:18  
AIR DRYING 9:10  
CYCLE COMPLETE 9:19

-----  
Digital Signature #  
62B4C19C24B74364  
-----