
Midmark M11 - v1.0.0
Total Cycles: 4266

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2024 04:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.2 F 0.0
2:00 165.1 F 3.3
4:00 193.5 F 7.4
6:00 207.2 F 12.6
8:00 223.4 F 6.7
10:00 238.0 F 11.6
12:00 234.5 F 8.4
14:00 242.6 F 11.2
16:00 245.8 F 12.9
18:00 257.7 F 19.4
20:00 268.8 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.3 F 28.6
2:00 271.3 F 28.3
3:00 272.2 F 28.5
4:00 271.1 F 28.2

Min 271.0 F 27.9
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 235.9 F 0.0
10:00 210.7 F 0.0
15:00 210.7 F 0.0
20:00 228.1 F 0.0
25:00 210.0 F 0.0
30:00 186.2 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

02/02/2024 05:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____