

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1133

Name: RITTER2  
Sterilizer ID: V2313228  
05/27/2022 02:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.2 F 0.0  
2:00 174.6 F 3.9  
4:00 195.4 F 9.1  
6:00 233.1 F 8.8  
8:00 232.5 F 11.4  
10:00 236.4 F 9.4  
12:00 236.2 F 9.2  
14:00 249.5 F 15.5  
16:00 262.1 F 22.6  
18:00 272.4 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 271.9 F 30.1  
2:00 271.9 F 29.6  
3:00 272.4 F 30.2  
4:00 271.4 F 29.8  
  
Min 270.3 F 28.8  
Max 273.4 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
5:00 233.4 F 0.0  
10:00 229.2 F 0.0  
15:00 229.3 F 0.0  
20:00 230.4 F 0.0  
25:00 229.5 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 2:20  
HEATING: 18:05  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:57:32  
CYCLE COMPLETE

05/27/2022 03:54 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_