
Midmark M11 - v1.0.5
Total Cycles: 2620

Name: RITTER2
Sterilizer ID: V2313228
07/06/2023 03:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.9 F 0.0
2:00 152.3 F 0.1
4:00 174.6 F 3.7
6:00 206.5 F 8.5
8:00 213.4 F 13.0
10:00 225.2 F 8.5
12:00 238.9 F 12.1
14:00 239.2 F 11.4
16:00 238.5 F 10.1
18:00 251.9 F 16.8
20:00 263.4 F 23.7
22:00 272.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.7 F 29.8
2:00 271.5 F 29.0
3:00 272.1 F 29.9
4:00 272.3 F 29.5

Min 271.3 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 235.1 F 0.0
10:00 229.9 F 0.0
15:00 229.5 F 0.0
20:00 230.4 F 0.0
25:00 226.8 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 22:09
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:35
CYCLE COMPLETE

07/06/2023 04:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____