
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006481
9:01 17/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:29
STERILIZING 3:23
136.0 C 220kPa 3:23
136.4 C 221kPa 3:53
136.4 C 225kPa 4:23
136.4 C 224kPa 4:53
135.9 C 221kPa 5:23
136.4 C 220kPa 5:53
136.4 C 224kPa 6:23
136.4 C 225kPa 6:53
136.4 C 225kPa 6:53
VENTING 6:53
AIR DRYING 7:42
CYCLE COMPLETE 13:38

Digital Signature #
DEC6E0C04B5DFC34
