
Midmark M11 - v1.0.0
Total Cycles: 3526

Name: RITTER 1
Sterilizer ID: V2178198
07/14/2023 08:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.7 F 0.0
2:00 72.8 F 0.0
4:00 81.7 F 0.0
6:00 110.0 F 0.1
8:00 144.3 F 1.7
10:00 171.3 F 3.9
12:00 193.2 F 6.7
14:00 203.9 F 10.5
16:00 216.4 F 6.9
18:00 224.0 F 7.5
20:00 233.7 F 11.5
22:00 231.3 F 7.5
24:00 243.9 F 13.0
26:00 241.4 F 11.0
28:00 253.5 F 17.1
30:00 263.9 F 23.3
32:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.8 F 29.2
2:00 271.9 F 28.4
3:00 271.8 F 28.8
4:00 272.5 F 29.0

Min 271.4 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 200.6 F 0.0
10:00 192.0 F 0.0
15:00 189.7 F 0.0
20:00 213.4 F 0.0
25:00 208.3 F 0.0
30:00 209.4 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:13
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:10:13
CYCLE COMPLETE

07/14/2023 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____