
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001943
8:49 12/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:40
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:40

Digital Signature #
A53F98EACFCFEC3A
