

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3203

Name: RITTER2  
Sterilizer ID: V2313228  
12/11/2023 12:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.0 F 0.0  
2:00 151.5 F 1.5  
4:00 170.1 F 4.5  
6:00 192.1 F 8.5  
8:00 209.2 F 12.9  
10:00 217.2 F 6.9  
12:00 229.4 F 11.7  
14:00 232.7 F 8.8  
16:00 239.8 F 10.6  
18:00 246.0 F 13.7  
20:00 257.6 F 20.1  
22:00 268.3 F 27.0  
24:00 271.8 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.5  
1:00 272.6 F 30.2  
2:00 271.7 F 29.6  
3:00 271.7 F 30.0  
4:00 272.6 F 30.2

Min 271.2 F 29.4  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 171.4 F 0.0  
5:00 199.5 F 0.0  
10:00 195.6 F 0.0  
15:00 192.3 F 0.0  
20:00 188.3 F 0.0  
25:00 186.3 F 0.0  
30:00 183.4 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 24:08  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:02:45  
CYCLE COMPLETE

12/11/2023 01:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_