
Midmark M11 - v1.0.0
Total Cycles: 1834

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2022 12:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.1 F 0.0
2:00 130.3 F 0.4
4:00 162.4 F 3.2
6:00 185.5 F 7.4
8:00 221.5 F 12.1
10:00 220.9 F 5.5
12:00 233.3 F 10.1
14:00 234.8 F 8.2
16:00 242.1 F 11.8
18:00 239.6 F 9.9
20:00 252.8 F 16.5
22:00 264.4 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.4 F 28.5
2:00 271.8 F 28.6
3:00 272.1 F 28.3
4:00 272.2 F 28.8

Min 271.2 F 28.0
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 223.8 F 0.0
10:00 208.2 F 0.0
15:00 211.9 F 0.0
20:00 210.0 F 0.0
25:00 236.6 F 0.0
30:00 241.6 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 23:55
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:56
CYCLE COMPLETE

04/18/2022 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____