

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3632

Name: RITTER 1  
Sterilizer ID: V2178198  
08/10/2023 08:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.6 F 0.0  
2:00 189.8 F 2.1  
4:00 191.5 F 6.5  
6:00 213.2 F 12.0  
8:00 222.3 F 6.6  
10:00 235.0 F 12.0  
12:00 235.9 F 9.4  
14:00 238.1 F 9.3  
16:00 248.1 F 14.4  
18:00 260.1 F 21.0  
20:00 271.0 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.2 F 28.9  
2:00 271.2 F 28.5  
3:00 272.2 F 28.9  
4:00 271.1 F 28.2  
  
Min 271.0 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 206.3 F 0.0  
10:00 210.0 F 0.0  
15:00 210.9 F 0.0  
20:00 210.1 F 0.0  
25:00 209.6 F 0.0  
30:00 210.1 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 20:43  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:59:46  
CYCLE COMPLETE

08/10/2023 09:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_