
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001604
15:56 16/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:02
STERILIZING 3:01
136.5 C 224kPa 3:01
136.3 C 224kPa 3:31
136.2 C 224kPa 4:01
136.4 C 225kPa 4:31
136.4 C 225kPa 5:01
136.2 C 223kPa 5:31
136.4 C 225kPa 6:01
136.4 C 224kPa 6:31
136.4 C 224kPa 6:31
VENTING 6:31
AIR DRYING 7:22
CYCLE COMPLETE 7:26

Digital Signature #
1CB9006133D0B868
