
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001604
16:00 11/19/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:13
PRESSURIZING 6:15
STERILIZING 7:53
136.0 C 220kPa 7:53
136.1 C 219kPa 8:23
136.2 C 220kPa 8:53
136.5 C 223kPa 9:23
136.4 C 219kPa 9:53
136.3 C 220kPa 10:24
136.4 C 222kPa 10:54
136.0 C 219kPa 11:23
VENTING 11:23
AIR DRYING 12:25
CYCLE COMPLETE 17:51

Digital Signature #
122D1A5383B08847
