
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 004217
14:43 23/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:43
PRESSURIZING 1:49
STERILIZING 2:48
136.4 C 221kPa 2:48
136.4 C 223kPa 3:19
136.2 C 222kPa 3:49
136.4 C 220kPa 4:19
136.0 C 223kPa 4:49
136.2 C 224kPa 5:19
136.4 C 224kPa 5:49
136.4 C 225kPa 6:19
136.4 C 225kPa 6:19
VENTING 6:19
AIR DRYING 7:03
CYCLE COMPLETE 9:49

Digital Signature #
C92ACA741D807910
