
Midmark M11 - v1.0.5
Total Cycles: 3889

Name: RITTER2
Sterilizer ID: V2313228
08/23/2024 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.2 F 0.0
2:00 152.0 F 1.8
4:00 163.5 F 4.8
6:00 195.1 F 8.7
8:00 204.4 F 12.8
10:00 217.9 F 6.3
12:00 228.1 F 10.5
14:00 231.9 F 8.4
16:00 240.4 F 12.1
18:00 239.1 F 10.6
20:00 251.1 F 16.4
22:00 261.5 F 22.4
24:00 271.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.7 F 29.9
2:00 271.9 F 29.3
3:00 271.1 F 29.2
4:00 272.4 F 29.9

Min 271.1 F 29.0
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 210.7 F 0.0
10:00 202.7 F 0.0
15:00 194.0 F 0.0
20:00 192.5 F 0.0
25:00 194.5 F 0.0
30:00 198.3 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 24:48
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:51
CYCLE COMPLETE

08/23/2024 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____