
Midmark M11 - v1.0.5
Total Cycles: 4001

Name: RITTER2
Sterilizer ID: V2313228
10/04/2024 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.2 F 0.0
2:00 157.1 F 2.0
4:00 163.3 F 4.9
6:00 192.9 F 8.6
8:00 204.3 F 13.2
10:00 215.1 F 6.6
12:00 225.7 F 11.1
14:00 230.0 F 7.4
16:00 242.2 F 13.0
18:00 241.5 F 11.6
20:00 253.2 F 17.5
22:00 263.7 F 23.8
24:00 272.2 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.0
1:00 272.6 F 29.8
2:00 273.3 F 30.1
3:00 272.2 F 29.6
4:00 271.7 F 29.6

Min 271.0 F 28.9
Max 273.4 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 211.7 F 0.0
10:00 215.7 F 0.0
15:00 213.9 F 0.0
20:00 201.8 F 0.0
25:00 199.9 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:16
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:06
CYCLE COMPLETE

10/04/2024 12:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____