
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000276
10:00 23/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:07
STERILIZING 4:11
136.4 C 223kPa 4:11
136.1 C 222kPa 4:41
136.3 C 223kPa 5:11
136.4 C 223kPa 5:41
136.2 C 222kPa 6:11
136.4 C 224kPa 6:41
136.2 C 220kPa 7:11
136.3 C 223kPa 7:41
136.3 C 223kPa 7:41
VENTING 7:41
AIR DRYING 8:31
CYCLE COMPLETE 8:34

Digital Signature #
BD0FC6CE6DB881CE
