
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001732
10:26 15/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:10
STERILIZING 3:13
136.4 C 220kPa 3:13
136.3 C 225kPa 3:43
136.4 C 224kPa 4:13
136.4 C 224kPa 4:43
136.4 C 225kPa 5:13
136.3 C 223kPa 5:43
136.4 C 225kPa 6:13
136.3 C 223kPa 6:43
136.3 C 223kPa 6:43
VENTING 6:44
AIR DRYING 7:34
CYCLE COMPLETE 8:30

Digital Signature #
ECD15F873E2DC7F0
