
Midmark M11 - v1.0.5
Total Cycles: 1284

Name: RITTER2
Sterilizer ID: V2313228
07/12/2022 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.6 F 0.0
2:00 82.5 F 0.0
4:00 104.6 F 0.0
6:00 120.4 F 0.1
8:00 156.4 F 0.2
10:00 189.0 F 0.6
12:00 204.7 F 4.5
14:00 216.5 F 8.0
16:00 231.7 F 11.8
18:00 229.8 F 7.4
20:00 241.8 F 12.5
22:00 239.5 F 10.5
24:00 236.5 F 9.2
26:00 251.1 F 16.1
28:00 262.9 F 22.9
30:00 272.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.3 F 28.9
2:00 272.9 F 29.5
3:00 271.9 F 29.0
4:00 272.3 F 28.9

Min 271.6 F 28.5
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.6 F 0.0
Dry Cycle Aborted

FILLING: 0:57
HEATING: 30:09
STERILIZING: 4:00
VENTING: 2:56
DRYING: 1:44
TOTAL CYCLE: 00:39:46

07/12/2022 01:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____