
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000569
14:23 06/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:55
STERILIZING 4:25
136.4 C 232kPa 4:25
136.2 C 231kPa 4:55
136.2 C 231kPa 5:25
136.3 C 231kPa 5:55
136.4 C 233kPa 6:25
136.3 C 232kPa 6:55
136.2 C 231kPa 7:25
136.4 C 231kPa 7:55
136.4 C 231kPa 7:55
VENTING 7:55
AIR DRYING 8:50
CYCLE COMPLETE 12:31

Digital Signature #
F82DC444D57A7B6D
