

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000386  
8:47 25/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:48  
PRESSURIZING 3:49  
STERILIZING 4:58  
136.4 C 232kPa 4:58  
135.9 C 229kPa 5:28  
136.2 C 231kPa 5:58  
136.4 C 233kPa 6:28  
136.4 C 233kPa 6:58  
136.4 C 232kPa 7:28  
136.3 C 232kPa 7:58  
136.2 C 231kPa 8:28  
136.2 C 231kPa 8:28  
VENTING 8:28  
AIR DRYING 9:24  
CYCLE COMPLETE 9:25

-----  
Digital Signature #  
E9255E8E5B9C53A5  
-----