
Midmark M11 - v1.0.0
Total Cycles: 4384

Name: RITTER 1
Sterilizer ID: V2178198
03/20/2024 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 71.3 F 0.0
4:00 101.8 F 0.6
6:00 108.7 F 2.5
8:00 133.4 F 5.1
10:00 158.4 F 8.2
12:00 180.7 F 11.4
14:00 197.9 F 2.5
16:00 210.2 F 5.4
18:00 220.3 F 8.6
20:00 230.7 F 12.4
22:00 230.2 F 7.9
24:00 241.8 F 12.8
26:00 239.8 F 10.7
28:00 251.8 F 16.6
30:00 262.3 F 22.6
32:00 272.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.4 F 29.0
2:00 272.5 F 29.5
3:00 271.6 F 28.7
4:00 272.5 F 29.2

Min 271.6 F 28.6
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 210.9 F 0.0
10:00 215.1 F 0.0
15:00 221.6 F 0.0
20:00 226.6 F 0.0
25:00 226.8 F 0.0
30:00 229.4 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:28
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:10:49
CYCLE COMPLETE

03/20/2024 09:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____