
Midmark M11 - v1.0.0
Total Cycles: 3999

Name: RITTER 1
Sterilizer ID: V2178198
11/09/2023 10:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.3 F 0.0
2:00 170.7 F 1.9
4:00 183.0 F 5.3
6:00 201.8 F 10.2
8:00 219.9 F 5.7
10:00 229.6 F 8.6
12:00 243.7 F 13.4
14:00 239.5 F 10.5
16:00 235.5 F 8.3
18:00 249.5 F 14.9
20:00 260.9 F 21.3
22:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.7
1:00 272.6 F 28.9
2:00 272.6 F 29.2
3:00 271.8 F 29.0
4:00 272.0 F 28.6

Min 271.4 F 28.6
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.8 F 0.0
5:00 206.7 F 0.0
10:00 196.7 F 0.0
15:00 195.0 F 0.0
20:00 193.1 F 0.0
25:00 189.9 F 0.0
30:00 188.9 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 23:12
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:02:16
CYCLE COMPLETE

11/09/2023 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____