
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004741
15:45 21/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:04
STERILIZING 3:10
136.4 C 222kPa 3:10
136.3 C 224kPa 3:40
136.4 C 224kPa 4:10
136.4 C 223kPa 4:40
136.2 C 222kPa 5:10
136.4 C 224kPa 5:40
136.2 C 224kPa 6:10
136.3 C 222kPa 6:40
136.3 C 222kPa 6:40
VENTING 6:40
AIR DRYING 7:29
CYCLE COMPLETE 9:05

Digital Signature #
40C88EEA72B83F1D
