
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002294
11:33 19/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 2:01
STERILIZING 3:10
136.4 C 223kPa 3:10
136.4 C 225kPa 3:40
136.3 C 225kPa 4:10
136.4 C 226kPa 4:40
136.2 C 223kPa 5:10
136.4 C 225kPa 5:40
136.4 C 226kPa 6:10
136.3 C 225kPa 6:40
136.3 C 225kPa 6:40
VENTING 6:40
AIR DRYING 7:30
CYCLE COMPLETE 11:50

Digital Signature #
EF961B72EC139F8B
