
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001581
11:27 11/12/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:27
PRESSURIZING 6:10
STERILIZING 7:39
136.0 C 220kPa 7:39
136.4 C 220kPa 8:10
136.2 C 221kPa 8:40
135.8 C 219kPa 9:10
136.4 C 221kPa 9:40
136.4 C 223kPa 10:10
136.3 C 221kPa 10:40
136.2 C 221kPa 11:09
VENTING 11:10
AIR DRYING 12:15
CYCLE COMPLETE 42:15

Digital Signature #
02E5E9DA65CE9605
