
Midmark M11 - v1.0.5
Total Cycles: 3952

Name: RITTER2
Sterilizer ID: V2313228
09/18/2024 12:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.8 F 0.0
2:00 169.9 F 1.8
4:00 164.8 F 5.0
6:00 187.7 F 8.8
8:00 214.8 F 13.3
10:00 222.9 F 5.8
12:00 233.5 F 9.8
14:00 235.2 F 10.0
16:00 240.9 F 11.7
18:00 239.4 F 10.3
20:00 251.6 F 16.4
22:00 262.3 F 22.6
24:00 272.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.3 F 29.3
2:00 271.1 F 28.7
3:00 272.3 F 29.3
4:00 271.3 F 28.7

Min 271.0 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 228.6 F 0.0
10:00 222.4 F 0.0
15:00 205.2 F 0.0
20:00 212.8 F 0.0
25:00 218.0 F 0.0
30:00 215.0 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:28
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:03:20
CYCLE COMPLETE

09/18/2024 01:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____