
Midmark M11 - v1.0.0
Total Cycles: 2495

Name: RITTER 1
Sterilizer ID: V2178198
10/26/2022 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.9 F 0.0
2:00 187.0 F 4.3
4:00 199.9 F 8.9
6:00 212.7 F 10.2
8:00 227.1 F 8.8
10:00 237.7 F 10.7
12:00 241.2 F 11.7
14:00 239.6 F 10.3
16:00 252.6 F 17.0
18:00 264.3 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.4 F 29.4
2:00 271.9 F 28.8
3:00 271.3 F 28.8
4:00 272.3 F 29.5

Min 271.1 F 28.5
Max 272.4 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 219.4 F 0.0
10:00 238.0 F 0.0
15:00 242.1 F 0.0
20:00 237.0 F 0.0
25:00 226.0 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:58
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

10/26/2022 04:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____