
Midmark M11 - v1.0.0
Total Cycles: 683

Name: MIDMARK
Sterilizer ID: V2178198
02/22/2021 07:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 76.2 F 0.0
4:00 92.7 F 0.1
6:00 93.8 F 0.1
8:00 124.1 F 0.2
10:00 163.2 F 0.3
12:00 191.3 F 2.9
14:00 223.2 F 5.8
16:00 230.3 F 9.0
18:00 238.0 F 12.5
20:00 233.5 F 7.7
22:00 241.7 F 12.4
24:00 238.1 F 9.4
26:00 239.5 F 9.5
28:00 247.1 F 13.4
30:00 258.4 F 19.6
32:00 268.5 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.3 F 28.3
2:00 272.1 F 28.6
3:00 271.6 F 27.9
4:00 272.5 F 28.6

Min 271.3 F 27.8
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 217.7 F 0.0
10:00 238.8 F 0.0
15:00 229.6 F 0.0
20:00 222.5 F 0.0
25:00 216.9 F 0.0
Dry Cycle Aborted

FILLING: 0:52
HEATING: 33:12
STERILIZING: 4:00
VENTING: 3:01
DRYING: 28:03
TOTAL CYCLE: 01:09:08

02/22/2021 08:46 AM
ACCEPTED: X
REJECTED:
Operator: U01