

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 001905  
14:19 26/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:52  
PRESSURIZING 1:59  
STERILIZING 3:03  
136.0 C 221kPa 3:03  
136.4 C 226kPa 3:33  
136.4 C 226kPa 4:03  
136.4 C 226kPa 4:33  
136.4 C 221kPa 5:03  
136.4 C 226kPa 5:33  
136.4 C 226kPa 6:03  
136.4 C 226kPa 6:33  
136.4 C 226kPa 6:33  
VENTING 6:33  
AIR DRYING 7:24  
CASSETTE REMOVED  
CYCLE COMPLETE 15:35  
-----

Digital Signature #  
761CD2E56825B40E  
-----