
Midmark M11 - v1.0.5
Total Cycles: 818

Name: RITTER2
Sterilizer ID: V2313228
02/22/2022 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.7 F 0.0
2:00 159.9 F 2.0
4:00 183.3 F 5.9
6:00 205.9 F 11.1
8:00 219.2 F 5.8
10:00 235.3 F 11.8
12:00 235.0 F 9.1
14:00 236.0 F 8.3
16:00 250.0 F 15.5
18:00 263.6 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.8 F 29.3
2:00 272.0 F 29.1
3:00 272.5 F 29.0
4:00 272.3 F 29.3

Min 271.8 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 166.8 F 0.0
Dry Cycle Aborted

FILLING: 1:16
HEATING: 19:41
STERILIZING: 4:00
VENTING: 2:23
DRYING: 0:47
TOTAL CYCLE: 00:28:07

02/22/2022 10:32 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____