
Midmark M11 - v1.0.0
Total Cycles: 2341

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2022 08:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.6 F 0.0
2:00 141.6 F 1.5
4:00 153.1 F 4.2
6:00 169.9 F 8.8
8:00 208.0 F 7.8
10:00 218.7 F 6.5
12:00 231.8 F 11.0
14:00 231.7 F 6.8
16:00 243.6 F 12.8
18:00 240.9 F 10.9
20:00 253.7 F 17.3
22:00 264.9 F 23.9
24:00 271.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.5 F 29.0
2:00 272.4 F 29.3
3:00 271.8 F 28.7
4:00 272.5 F 29.3

Min 271.4 F 28.6
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.8 F 0.0
5:00 214.7 F 0.0
10:00 209.3 F 0.0
15:00 214.0 F 0.0
20:00 218.1 F 0.0
25:00 215.8 F 0.0
30:00 219.5 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:36
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:02:55
CYCLE COMPLETE

09/20/2022 09:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____