
Midmark M11 - v1.0.0
Total Cycles: 2622

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2022 04:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.8 F 0.0
2:00 167.7 F 3.5
4:00 181.3 F 7.6
6:00 199.2 F 12.3
8:00 215.6 F 6.1
10:00 229.7 F 11.1
12:00 230.6 F 7.6
14:00 243.3 F 13.6
16:00 243.4 F 12.1
18:00 255.9 F 18.7
20:00 267.2 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.2
2:00 271.7 F 28.5
3:00 271.9 F 29.1
4:00 272.4 F 29.1

Min 271.3 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 212.9 F 0.0
10:00 236.4 F 0.0
15:00 235.0 F 0.0
20:00 226.6 F 0.0
25:00 223.4 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:40
CYCLE COMPLETE

11/29/2022 05:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____