
Midmark M11 - v1.0.5
Total Cycles: 3718

Name: RITTER2
Sterilizer ID: V2313228
06/19/2024 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.4 F 0.0
2:00 150.0 F 2.2
4:00 168.8 F 5.4
6:00 191.0 F 9.2
8:00 217.3 F 13.5
10:00 222.4 F 6.6
12:00 232.8 F 10.8
14:00 231.8 F 7.4
16:00 242.5 F 12.6
18:00 241.1 F 11.2
20:00 252.9 F 17.1
22:00 263.5 F 23.4
24:00 272.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.8 F 29.5
2:00 271.9 F 28.9
3:00 271.2 F 28.9
4:00 272.7 F 29.6

Min 271.1 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 236.2 F 0.0
10:00 216.5 F 0.0
15:00 200.2 F 0.0
20:00 188.4 F 0.0
25:00 197.0 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:19
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:03:09
CYCLE COMPLETE

06/19/2024 02:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____