
Midmark M11 - v1.0.0
Total Cycles: 1428

Name: RITTER 1
Sterilizer ID: V2178198
12/07/2021 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.7 F 0.0
2:00 86.3 F 0.0
4:00 120.4 F 0.0
6:00 152.7 F 0.6
8:00 182.0 F 3.1
10:00 201.9 F 6.4
12:00 215.4 F 10.3
14:00 221.3 F 7.0
16:00 229.4 F 8.0
18:00 241.0 F 12.8
20:00 237.5 F 9.5
22:00 235.7 F 7.9
24:00 249.4 F 14.7
26:00 262.7 F 22.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.6 F 28.7
2:00 272.3 F 28.2
3:00 271.7 F 27.8
4:00 271.9 F 28.3

Min 271.6 F 27.8
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 219.8 F 0.0
5:00 223.4 F 0.0
10:00 156.3 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 27:55
STERILIZING: 4:00
VENTING: 2:30
DRYING: 14:51
TOTAL CYCLE: 00:50:24

12/07/2021 12:08 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____