
Midmark M11 - v1.0.5
Total Cycles: 632

Name: RITTER2
Sterilizer ID: V2313228
12/14/2021 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.9 F 0.0
2:00 177.0 F 3.3
4:00 185.8 F 7.8
6:00 205.9 F 13.1
8:00 221.7 F 7.3
10:00 234.7 F 12.8
12:00 237.9 F 10.6
14:00 237.4 F 9.4
16:00 251.5 F 16.5
18:00 263.5 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.2 F 29.6
2:00 272.6 F 29.3
3:00 271.5 F 28.9
4:00 272.6 F 29.8

Min 271.5 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.2 F 0.0
5:00 202.4 F 0.0
10:00 214.5 F 0.0
15:00 215.9 F 0.0
20:00 214.3 F 0.0
25:00 213.0 F 0.0
30:00 211.1 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:56
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

12/14/2021 01:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____