
Midmark M11 - v1.0.0
Total Cycles: 2656

Name: RITTER 1
Sterilizer ID: V2178198
12/07/2022 05:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.7 F 0.0
2:00 188.0 F 3.9
4:00 196.9 F 8.6
6:00 215.0 F 10.8
8:00 228.2 F 8.6
10:00 240.0 F 10.8
12:00 241.2 F 11.8
14:00 240.3 F 10.6
16:00 254.2 F 17.6
18:00 266.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 28.9
2:00 272.0 F 28.8
3:00 272.4 F 28.9
4:00 271.7 F 28.5

Min 271.6 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 228.5 F 0.0
10:00 236.6 F 0.0
15:00 231.3 F 0.0
20:00 226.8 F 0.0
25:00 224.2 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

12/07/2022 06:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____