

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1148

Name: RITTER 1  
Sterilizer ID: V2178198  
09/08/2021 10:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.9 F 0.0  
2:00 174.6 F 3.6  
4:00 188.9 F 7.9  
6:00 211.7 F 13.0  
8:00 222.2 F 9.0  
10:00 232.8 F 9.6  
12:00 239.1 F 12.4  
14:00 239.4 F 11.0  
16:00 253.6 F 18.0  
18:00 265.9 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.4 F 29.5  
2:00 272.1 F 29.6  
3:00 272.3 F 29.4  
4:00 272.2 F 29.6  
  
Min 271.6 F 29.0  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.3 F 0.0  
5:00 223.1 F 0.0  
10:00 225.5 F 0.0  
15:00 226.7 F 0.0  
20:00 200.2 F 0.0  
25:00 188.8 F 0.0  
Dry Cycle Aborted

FILLING: 1:43  
HEATING: 19:33  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 25:30  
TOTAL CYCLE: 00:53:56

09/08/2021 11:10 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01