
Midmark M11 - v1.0.5
Total Cycles: 4092

Name: RITTER2
Sterilizer ID: V2313228
11/06/2024 12:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.1 F 0.0
2:00 97.3 F 0.0
4:00 102.3 F 0.0
6:00 137.6 F 0.2
8:00 164.3 F 2.7
10:00 183.4 F 5.7
12:00 206.4 F 9.2
14:00 218.6 F 13.3
16:00 223.8 F 7.8
18:00 235.1 F 12.2
20:00 234.2 F 9.3
22:00 239.0 F 10.3
24:00 245.9 F 13.7
26:00 257.1 F 19.9
28:00 267.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.7 F 29.9
2:00 271.6 F 29.1
3:00 271.6 F 29.6
4:00 272.5 F 29.7

Min 271.2 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 191.1 F 0.0
10:00 199.7 F 0.0
15:00 196.5 F 0.0
20:00 196.7 F 0.0
25:00 197.2 F 0.0
30:00 196.3 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:29
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:08:03
CYCLE COMPLETE

11/06/2024 01:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____