

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
2.5uS / 1.6ppm  
CYCLE NUMBER 000031  
13:16 30/09/2019  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:17  
PRESSURIZING 2:19  
STERILIZING 3:17  
136.4 C 226kPa 3:17  
136.2 C 227kPa 3:47  
136.3 C 225kPa 4:17  
136.4 C 224kPa 4:47  
136.4 C 226kPa 5:17  
136.3 C 225kPa 5:47  
136.1 C 223kPa 6:17  
136.4 C 225kPa 6:47  
136.4 C 225kPa 6:47  
VENTING 6:47  
AIR DRYING 7:34  
CYCLE COMPLETE 35:53  
-----

Digital Signature #  
4A0A6C621361E0D8  
-----