
Midmark M11 - v1.0.0
Total Cycles: 1105

Name: RITTER 1
Sterilizer ID: V2178198
08/20/2021 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.2 F 0.0
2:00 119.3 F 0.0
4:00 139.5 F 0.1
6:00 169.0 F 2.0
8:00 193.6 F 6.5
10:00 209.7 F 11.0
12:00 227.1 F 5.3
14:00 234.4 F 10.8
16:00 233.3 F 7.7
18:00 244.3 F 11.9
20:00 244.8 F 12.5
22:00 257.3 F 19.1
24:00 268.4 F 26.1
26:00 271.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.5 F 29.3
2:00 271.9 F 29.0
3:00 272.4 F 29.2
4:00 271.8 F 29.0

Min 271.7 F 28.7
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.2 F 0.0
5:00 226.0 F 0.0
10:00 231.7 F 0.0
15:00 231.2 F 0.0
20:00 230.4 F 0.0
25:00 227.1 F 0.0
Dry Cycle Aborted

FILLING: 0:52
HEATING: 26:06
STERILIZING: 4:00
VENTING: 2:59
DRYING: 26:02
TOTAL CYCLE: 00:59:58

08/20/2021 12:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____