

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5016

Name: RITTER 1  
Sterilizer ID: V2178198  
10/23/2024 09:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.0 F 0.0  
2:00 116.8 F 0.0  
4:00 123.6 F 0.6  
6:00 154.2 F 0.8  
8:00 179.2 F 2.8  
10:00 207.4 F 6.2  
12:00 215.2 F 9.7  
14:00 225.8 F 13.5  
16:00 226.8 F 8.0  
18:00 237.6 F 12.5  
20:00 237.0 F 9.6  
22:00 239.3 F 9.7  
24:00 246.5 F 13.5  
26:00 257.7 F 19.4  
28:00 267.7 F 25.6  
30:00 272.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.7  
1:00 272.6 F 29.1  
2:00 271.9 F 28.9  
3:00 272.4 F 29.0  
4:00 272.0 F 29.0

Min 271.7 F 28.6  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.8 F 0.0  
5:00 212.8 F 0.0  
10:00 203.2 F 0.0  
15:00 208.7 F 0.0  
20:00 218.9 F 0.0  
25:00 232.4 F 0.0  
30:00 233.6 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 30:13  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:08:47  
CYCLE COMPLETE

10/23/2024 11:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_