
Midmark M11 - v1.0.0
Total Cycles: 5075

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2024 03:53 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 81.5 F 0.0
2:00 84.9 F 0.0
4:00 152.4 F 0.0
6:00 175.7 F 0.3

C103: HEATUP MODE
STOP PRESSED

** ITEMS NOT STERILE **

11/08/2024 04:01 PM