
Midmark M11 - v1.0.5
Total Cycles: 4022

Name: RITTER2
Sterilizer ID: V2313228
10/11/2024 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 87.7 F 0.0
4:00 78.6 F 0.0
6:00 107.3 F 0.1
8:00 149.5 F 0.3
10:00 178.3 F 0.5
12:00 196.7 F 3.6
14:00 224.8 F 7.0
16:00 231.4 F 10.5
18:00 228.3 F 8.1
20:00 234.8 F 9.6
22:00 238.8 F 10.4
24:00 243.2 F 12.3
26:00 241.0 F 11.0
28:00 253.6 F 17.4
30:00 264.4 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.5 F 29.0
2:00 272.2 F 29.2
3:00 271.9 F 28.6
4:00 272.7 F 29.4

Min 271.6 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 230.2 F 0.0
10:00 191.3 F 0.0
15:00 190.4 F 0.0
20:00 190.2 F 0.0
25:00 193.3 F 0.0
30:00 198.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:60
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:10:07
CYCLE COMPLETE

10/11/2024 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____