
Midmark M11 - v1.0.5
Total Cycles: 3773

Name: RITTER2
Sterilizer ID: V2313228
07/11/2024 06:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.0 F 0.0
2:00 162.9 F 2.9
4:00 188.0 F 7.8
6:00 211.3 F 8.9
8:00 230.0 F 9.0
10:00 235.3 F 9.5
12:00 244.0 F 13.0
14:00 243.7 F 12.3
16:00 257.8 F 19.7
18:00 269.9 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.2 F 28.7
2:00 272.7 F 29.3
3:00 272.0 F 29.1
4:00 272.1 F 28.8

Min 271.7 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.4 F 0.0
5:00 232.9 F 0.0
10:00 204.2 F 0.0
15:00 204.1 F 0.0
20:00 212.1 F 0.0
25:00 218.6 F 0.0
30:00 220.6 F 0.0

Drying Complete

FILLING: 1:60
HEATING: 18:50
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

07/11/2024 07:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____