
Midmark M11 - v1.0.0
Total Cycles: 3108

Name: RITTER 1
Sterilizer ID: V2178198
04/13/2023 01:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.4 F 0.0
2:00 188.5 F 3.6
4:00 208.5 F 8.6
6:00 224.4 F 8.7
8:00 234.0 F 9.9
10:00 236.3 F 8.4
12:00 245.8 F 13.6
14:00 244.6 F 12.5
16:00 257.4 F 19.3
18:00 268.5 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 28.9
2:00 271.4 F 28.2
3:00 272.5 F 29.1
4:00 271.9 F 28.5

Min 271.3 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 224.8 F 0.0
10:00 230.1 F 0.0
15:00 230.0 F 0.0
20:00 228.2 F 0.0
25:00 229.4 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

04/13/2023 02:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____